



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST-NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/783,895	02/20/2004	Roger P. Jackson	10,187	7627
<div>7590 John C. McMahon PO Box 30069 Kansas City, MO 64112</div>				
			EXAMINER COMSTOCK, DAVID C	
			ART UNIT 3733	PAPER NUMBER
			MAIL DATE 07/19/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/783,895

Applicant(s)

JACKSON, ROGER P.

Examiner

David Comstock

Art Unit

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 23 April 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-21 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-21 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 20 February 2004 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-21 are rejected under 35 U.S.C. 103(a) as being unpatentable over DE 298 10 798 U1 (Shafer) (cited by applicant) in view of US 6,056,753 (Jackson) (applicant's prior patent).

Shafer discloses a threaded bone screw device 1 comprising an open head formed by a pair of spaced apart arms 4 and a cylindrical closure 3 (see, e.g., Figs. 1 and 2). The arms define a receiving channel for a bone fixation rod 2. Facing surfaces of the arms define guide and advancement structure 9 for rotatable mating with a guide flange on the closure. The guide flange is continuous and helical and has a compound contour including an enlarged outer periphery with an inward facing component. The leading end comprises a V-shaped set ring 15. Shafer does not disclose a closure comprising a multi-stepped bore, an installation head with a breakaway region or a set point. Jackson discloses providing a closure 1 for a similar device wherein the closure comprises a multi-stepped bore defining shoulders 41-43, an installation head 20, a breakaway region 24, and a set point 51 on the leading end (see, e.g., Figs. 3 and 5 and col. 3, lines 22-65, col. 5, lines 53-65, and col. 7, line 61 - col. 8, line 10). The multi-

Art Unit: 3733

stepped bore facilitates the use of an easy out tool for removal of the closure. The installation head and breakaway region allow the closure to be tightened to an optimal predetermined torque and prevent overtightening. The set point penetrates into the rod to prevent relative motion and better secure the system. Accordingly, it would have been obvious to one having ordinary skill in the art at the time the invention was made to have provided the device of Shafer with a closure comprising a multi-stepped bore, an installation head with a breakaway region, and a set point, in view of Jackson, in order to allow the use of an easy out tool for removal of the closure, to allow the closure to be tightened to an optimal predetermined torque and prevent overtightening, and to allow the leading end of the closure to penetrate into the rod to prevent relative motion and better secure the system.

Response to Arguments

Applicant's arguments filed 23 April 2007 have been fully considered but they are not persuasive.

Shafer discloses a cylindrical closure with a guide and advancement flange extending helically about an outer cylindrical structure, as set forth in the rejection. In addition, it is noted that the flange comprises at least a portion that is continuous.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to David Comstock whose telephone number is (571) 272-4710. Please leave a detailed voice message if examiner is unavailable. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached at (571) 272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3733

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



D. Comstock



EDUARDO C. ROBERT
SUPERVISORY PATENT EXAMINER